

FROM POLICY TO PRACTICE: OSAP'S GUIDE TO THE CDC GUIDELINES

A Step-By-Step
Dental Infection
Prevention and Control
Implementation
Workbook

Updated 2022



From Policy to Practice: OSAP's Guide to the CDC Guidelines

Your tool for applying CDC dental infection prevention and control guidelines



An education and training resource
prepared for dental healthcare personnel by OSAP —
the Organization for Safety, Asepsis and Prevention

From Policy to Practice: OSAP's Guide to the CDC Guidelines is an education and training tool produced by the Organization for Safety, Asepsis and Prevention (OSAP) and supported by Cooperative Agreement No. U58/CCU318566-02 from the U.S. Centers for Disease Control and Prevention. Its contents are solely the responsibility of OSAP and do not necessarily represent the official views of the Centers for Disease Control and Prevention.

Copyright © 2004 - 2022 by OSAP.

All rights reserved. No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, recording, or otherwise, without written permission of OSAP.

Printed in the United States of America

OSAP
One Glenlake Parkway, NE
Suite 1200
Atlanta, GA 30328

ISBN: 978-0-9752519-4-2

IMPORTANT DISCLAIMER: Although OSAP believes that the information contained herein is accurate, it necessarily reflects OSAP's interpretation of CDC guidelines. Moreover, inadvertent errors may occur. Accordingly, OSAP makes no representations of any kind that its interpretations are always correct, complete, or up-to-date and expressly disclaims any representation that this workbook satisfies any applicable standard of care. Users of this workbook are encouraged to read the Centers for Disease Control and Prevention guidelines and reach their own conclusions regarding any matter subject to interpretation. OSAP shall not be liable for any direct, indirect, incidental, special, or consequential damages resulting from the user's reliance upon the material contained herein.

- Develop and update standard operating procedures based on manufacturer instructions and train staff members on waterline maintenance and monitoring tasks.
- Keep records of monitoring results along with other quality assurance records such as the sterilizer monitoring log. These records should include the unit or room, date and time of sample collection, the type of test performed (in-office, laboratory), results, and methods used to correct problems including removal from service if needed.

If bacterial counts exceed 500 CFU/mL, evaluate for possible technique errors, re-treat the dental unit water, and retest. If initial testing used a pooled sample, consider testing individual lines and source water to identify the source of contamination. If a unit fails two consecutive tests, consider removing it from service and contact the manufacturer. Continue to monitor, evaluate techniques, and treat waterlines until acceptable water quality is regularly attained.

Boil-Water Advisory Dos and Don'ts

During a Boil-Water Advisory

DO ...

- ... Have patients use bottled water or distilled water to rinse.
- ... Use antimicrobial products that do not require water, such as alcohol-based hand rubs, for hand hygiene. If hands are visibly soiled, use bottled water and soap for handwashing or an antiseptic-containing towelette.

DON'T ...

- ... Use water from the public water system for dental treatment, patient rinsing, or handwashing.
- ... Use tap water to dilute germicides or for hand hygiene (unless the water has been brought to a rolling boil for at least 1 minute and cooled before use).

After the boil-water advisory is lifted...

DO ...

- ... Follow the local water utility's guidance for flushing all waterlines served by the public water system.
- ... Disinfect dental operative waterlines according to the manufacturer's instructions.